



**National Accreditation Board for
Testing and Calibration Laboratories**

(A Constituent Board of Quality Council of India)



CERTIFICATE OF ACCREDITATION

HITECH INDIA EQUIPMENTS PVT. LTD.

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2005

"General Requirements for the Competence of Testing & Calibration Laboratories"

for its facilities at

New No. 26, Old No. 91, Karpagam Avenue , 3rd Street, R.A. Puram,
Chennai, Tamil Nadu

in the field of

CALIBRATION

Certificate Number CC-2582 (in lieu of C-0264)

Issue Date 14/02/2018

Valid Until 13/02/2020

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.

(To see the scope of accreditation of this laboratory, you may also visit NABL website www.nabl-india.org)

Signed for and on behalf of NABL

Avijit Das
Program Director



89076970200020000306

Anil Relia
Chief Executive Officer



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Hitech India Equipments Pvt. Ltd., New No. 26, Old No. 91, Karpagam Avenue, 3rd Street, R.A. Puram, Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2582 (In lieu of C-0264)

Page 1 of 2

Validity 14.02.2018 to 13.02.2020

Last Amended on 20.02.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
<u>MECHANICAL CALIBRATION</u>				
I. UTM, TENSION CREEP AND TORSION TESTING MACHINE				
1.	Force *	1 N to 50 kN(Tension)	0.57%	Using Poving Ring & Load Cell as per IS 1828
	Verification of UTM/CTM/TTM	1 kN to 1000 kN(Compression)	0.30%	
II. IMPACT TESTING MACHINE				
2.	Impact *			Using Clinometer, Spirit Level, Load Cell by direct method as per ISO 148
	Charpy	0 to 300 J	1.0%	
	Izod	0 to 170 J	1.3%	
III. HARDNESS TESTING MACHINES				
1.	Verification of Rockwell Hardness Tester *	HRA HRB HRC	1.62 HRA 1.51 HRB 1.16 HRC	Using Standard Test Blocks by Indirect Method as per IS 1586
2.	Verification of Rockwell Superficial Hardness Tester *	HR 15N HR 30N HR 45N HR 15T HR 30T HR 45T	1 HR 15N 1.76 HR 30N 1.48 HR 45N 1.79 HR 15T 1 HR 30T 1 HR 45T	Using Standard Test Blocks by Indirect Method as per IS 1586

Dheeraj Chawla
Convenor

Avijit Das
Program Director



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Hitech India Equipments Pvt. Ltd., New No. 26, Old No. 91, Karpagam Avenue, 3rd Street, R.A. Puram, Chennai, Tamil Nadu

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2582 (In lieu of C-0264) **Page** 2 of 2

Validity 14.02.2018 to 13.02.2020 **Last Amended on** 20.02.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (\pm)	Remarks
3.	Verification of Vickers Hardness Tester *	HV 1 HV 5 HV 10 HV 30	2% 2.22% 2.00% 2.00%	Using Standard Test Blocks by Indirect Method as per IS 1501
4.	Verification of Brinell Hardness Tester *	HBW 2.5/187.5 HBW 5/250 HBW 5/750 HBW 10/3000	2% 1.8% 2.41% 2.00%	Using Standard Test Blocks by Indirect Method as per IS 1500

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

*Only for Site Calibration

Dheeraj Chawla
Convenor

Avijit Das
Program Director